

## IV

(Notices)

## NOTICES FROM EUROPEAN UNION INSTITUTIONS, BODIES, OFFICES AND AGENCIES

## EUROPEAN COMMISSION

**Commission communication in the framework of the implementation of Council Directive 90/385/EEC of 20 June 1990 on the approximation of the laws of the Member States relating to active implantable medical devices**

(Text with EEA relevance)

(Publication of titles and references of harmonised standards under the directive)

(2011/C 16/01)

| ESO (1) | Reference and title of the harmonised standard (and reference document)  | First publication OJ          | Reference of superseded standard  | Date of cessation of presumption of conformity of superseded standard<br>Note 1 |
|---------|--|-------------------------------|-----------------------------------|---|
| CEN     | EN 556-1:2001<br>Sterilisation of medical devices — Requirements for medical devices to be designated 'Sterile' — Part 1: Requirements for terminally sterilised medical devices | 31.7.2002                     | EN 556:1994 + A1:1998<br>Note 2.1 | Date expired<br>(30.4.2002)   |
|         | EN 556-1:2001/AC:2006  | 15.11.2006                    |                                   |   |
| CEN     | EN 556-2:2003<br>Sterilisation of medical devices — Requirements for medical devices to be designated 'Sterile' — Part 2: Requirements for aseptically processed medical devices | 9.8.2007                      |                                   |   |
| CEN     | EN 980:2008<br>Symbols for use in the labelling of medical devices   | 23.7.2008                     | EN 980:2003<br>Note 2.1           | Date expired<br>(31.5.2010)   |
| CEN     | EN 1041:2008<br>Information supplied by the manufacturer of medical devices  | 19.2.2009                     | EN 1041:1998<br>Note 2.1          | 31.8.2011   |
| CEN     | EN ISO 10993-1:2009<br>Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009)                             | 2.12.2009                     | EN ISO 10993-1:2009<br>Note 2.1   | Date expired<br>(21.3.2010)   |
|         | EN ISO 10993-1:2009/AC:2010  | This is the first publication |                                   |   |

| ESO <sup>(1)</sup> | Reference and title of the harmonised standard (and reference document)  | First publication OJ          | Reference of superseded standard | Date of cessation of presumption of conformity of superseded standard<br>Note 1 |
|--------------------|--|-------------------------------|----------------------------------|---|
| CEN                | EN ISO 10993-4:2009<br>Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood (ISO 10993-4:2002, including Amd 1:2006)                        | 2.12.2009                     | EN ISO 10993-4:2002<br>Note 2.1  | Date expired (21.3.2010)  |
| CEN                | EN ISO 10993-5:2009<br>Biological evaluation of medical devices — Part 5: Tests for <i>in vitro</i> cytotoxicity (ISO 10993-5:2009)  | 2.12.2009                     | EN ISO 10993-5:1999<br>Note 2.1  | Date expired (31.12.2009)   |
| CEN                | EN ISO 10993-6:2009<br>Biological evaluation of medical devices — Part 6: Tests for local effects after implantation (ISO 10993-6:2007)  | 2.12.2009                     | EN ISO 10993-6:2007<br>Note 2.1  | Date expired (21.3.2010)  |
| CEN                | EN ISO 10993-7:2008<br>Biological evaluation of medical devices — Part 7: Ethylene oxide sterilisation residuals (ISO 10993-7:2008)  | 7.7.2010                      |                                  |   |
|                    | EN ISO 10993-7:2008/AC:2009  | 7.7.2010                      |                                  |   |
| CEN                | EN ISO 10993-9:2009<br>Biological evaluation of medical devices — Part 9: Framework for identification and quantification of potential degradation products (ISO 10993-9:2009)           | 2.12.2009                     | EN ISO 10993-9:2009<br>Note 2.1  | Date expired (21.3.2010)  |
| CEN                | EN ISO 10993-11:2009<br>Biological evaluation of medical devices — Part 11: Tests for systemic toxicity (ISO 10993-11:2006)  | 2.12.2009                     | EN ISO 10993-11:2006<br>Note 2.1 | Date expired (21.3.2010)  |
| CEN                | EN ISO 10993-12:2009<br>Biological evaluation of medical devices — Part 12: Sample preparation and reference materials (ISO 10993-12:2007)   | 2.12.2009                     | EN ISO 10993-12:2007<br>Note 2.1 | Date expired (21.3.2010)  |
| CEN                | EN ISO 10993-13:2010<br>Biological evaluation of medical devices — Part 13: Identification and quantification of degradation products from polymeric medical devices (ISO 10993-13:2010) | This is the first publication | EN ISO 10993-13:2009<br>Note 2.1 | Date expired (31.12.2010)   |
| CEN                | EN ISO 10993-16:2010<br>Biological evaluation of medical devices — Part 16: Toxico-kinetic study design for degradation products and leachables (ISO 10993-16:2010)                      | 7.7.2010                      | EN ISO 10993-16:2009<br>Note 2.1 | Date expired (31.8.2010)  |
| CEN                | EN ISO 10993-17:2009<br>Biological evaluation of medical devices — Part 17: Establishment of allowable limits for leachable substances (ISO 10993-17:2002)                               | 2.12.2009                     | EN ISO 10993-17:2002<br>Note 2.1 | Date expired (21.3.2010)  |

| ESO <sup>(1)</sup> | Reference and title of the harmonised standard (and reference document)  | First publication OJ | Reference of superseded standard                               | Date of cessation of presumption of conformity of superseded standard<br>Note 1 |
|--------------------|--|----------------------|--|---|
| CEN                | EN ISO 10993-18:2009<br>Biological evaluation of medical devices — Part 18: Chemical characterisation of materials (ISO 10993-18:2005)   | 2.12.2009            | EN ISO 10993-18:2005<br>Note 2.1                               | Date expired (21.3.2010)  |
| CEN                | EN ISO 11135-1:2007<br>Sterilisation of health-care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilisation process for medical devices (ISO 11135-1:2007) | 9.8.2007             | EN 550:1994<br>Note 2.1  | Date expired (31.5.2010)  |
| CEN                | EN ISO 11137-1:2006<br>Sterilisation of health-care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilisation process for medical devices (ISO 11137-1:2006)      | 7.9.2006             | EN 552:1994<br>Note 2.1  | Date expired (30.4.2009)  |
| CEN                | EN ISO 11137-2:2007<br>Sterilisation of health-care products — Radiation — Part 2: Establishing the sterilisation dose (ISO 11137-2:2006, corrected version 2006-08-01)  | 9.8.2007             |  |   |
|                    | EN ISO 11137-2:2007/AC:2009  | 2.12.2009            |  |   |
| CEN                | EN ISO 11138-2:2009<br>Sterilisation of health-care products — Biological indicators — Part 2: Biological indicators for ethylene oxide sterilisation processes (ISO 11138-2:2006)                                     | 2.12.2009            | EN ISO 11138-2:2006<br>Note 2.1                                | Date expired (21.3.2010)  |
| CEN                | EN ISO 11138-3:2009<br>Sterilisation of health-care products — Biological indicators — Part 3: Biological indicators for moist heat sterilisation processes (ISO 11138-3:2006)   | 2.12.2009            | EN ISO 11138-3:2006<br>Note 2.1                                | Date expired (21.3.2010)  |
| CEN                | EN ISO 11140-1:2009<br>Sterilisation of health-care products — Chemical indicators — Part 1: General requirements (ISO 11140-1:2005)   | 2.12.2009            | EN ISO 11140-1:2005<br>Note 2.1                                | Date expired (21.3.2010)  |
| CEN                | EN ISO 11607-1:2009<br>Packaging for terminally sterilised medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2006)                                      | 2.12.2009            | EN ISO 11607-1:2006<br>Note 2.1                                | Date expired (21.3.2010)  |
| CEN                | EN ISO 11737-1:2006<br>Sterilisation of medical devices — Microbiological methods — Part 1: Determination of a population of micro-organisms on products (ISO 11737-1:2006)  | 7.9.2006             | EN 1174-2:1996<br>EN 1174-1:1996<br>EN 1174-3:1996<br>Note 2.1 | Date expired (31.10.2006)   |
|                    | EN ISO 11737-1:2006/AC:2009  | 2.12.2009            |  |   |

| ESO <sup>(1)</sup> | Reference and title of the harmonised standard (and reference document)   | First publication OJ | Reference of superseded standard                   | Date of cessation of presumption of conformity of superseded standard<br>Note 1 |
|--------------------|---|----------------------|--|---|
| CEN                | EN ISO 11737-2:2009<br>Sterilisation of medical devices — Microbiological methods — Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilisation process (ISO 11737-2:2009)                              | 7.7.2010             |  |   |
| CEN                | EN ISO 13485:2003<br>Medical devices — Quality management systems — Requirements for regulatory purposes (ISO 13485:2003)   | 2.4.2004             | EN ISO 13488:2000<br>EN ISO 13485:2000<br>Note 2.1 | Date expired<br>(31.7.2009)   |
|                    | EN ISO 13485:2003/AC:2009   | 7.7.2010             |  |   |
| CEN                | EN 13824:2004<br>Sterilisation of medical devices — Aseptic processing of liquid medical devices — Requirements   | 24.6.2005            |  |   |
| CEN                | EN ISO 14155-1:2009<br>Clinical investigation of medical devices for human subjects — Part 1: General requirements (ISO 14155-1:2003)   | 7.7.2010             | EN ISO 14155-1:2003<br>Note 2.1                    | Date expired<br>(21.3.2010)   |
| CEN                | EN ISO 14155-2:2009<br>Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans (ISO 14155-2:2003)   | 7.7.2010             | EN ISO 14155-2:2003<br>Note 2.1                    | Date expired<br>(21.3.2010)   |
| CEN                | EN ISO 14937:2009<br>Sterilisation of health-care products — General requirements for characterisation of a sterilising agent and the development, validation and routine control of a sterilisation process for medical devices (ISO 14937:2009) | 7.7.2010             | EN ISO 14937:2000<br>Note 2.1                      | Date expired<br>(21.3.2010)   |
| CEN                | EN ISO 14971:2009<br>Medical devices — Application of risk management to medical devices (ISO 14971:2007, Corrected version 2007-10-01)   | 7.7.2010             | EN ISO 14971:2007<br>Note 2.1                      | Date expired<br>(21.3.2010)   |
| CEN                | EN ISO 17665-1:2006<br>Sterilisation of health-care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilisation process for medical devices (ISO 17665-1:2006)                            | 15.11.2006           | EN 554:1994<br>Note 2.1                            | Date expired<br>(31.8.2009)   |
| CEN                | EN 45502-1:1997<br>Active implantable medical devices — Part 1: General requirements for safety, marking and information to be provided by the manufacturer   | 27.8.1998            |  |   |
| CEN                | EN 45502-2-1:2004<br>Active implantable medical devices — Part 2-1: Particular requirements for active implantable medical devices intended to treat bradyarrhythmia (cardiac pacemakers)   | 24.6.2005            |  |   |

| ESO <sup>(1)</sup> | Reference and title of the harmonised standard (and reference document)  | First publication OJ          | Reference of superseded standard               | Date of cessation of presumption of conformity of superseded standard<br>Note 1 |
|--------------------|--|-------------------------------|--|---|
| CEN                | EN 45502-2-3:2010<br>Active implantable medical devices — Part 2-3: Particular requirements for cochlear and auditory brainstem implant systems  | 7.7.2010                      |  |   |
| Cenelec            | EN 45502-1:1997<br>Active implantable medical devices — Part 1: General requirements for safety, marking and information to be provided by the manufacturer (*)  | 27.8.1998                     |  |   |
| Cenelec            | EN 45502-2-1:2003<br>Active implantable medical devices — Part 2-1: Particular requirements for active implantable medical devices intended to treat bradyarrhythmia (cardiac pacemakers) (*)                  | 8.7.2004                      |  |   |
| Cenelec            | EN 45502-2-2:2008<br>Active implantable medical devices — Part 2-2: Particular requirements for active implantable medical devices intended to treat tachyarrhythmia (includes implantable defibrillators) (*) | 27.11.2008                    |  |   |
| Cenelec            | EN 45502-2-2:2008/AC:2009  | This is the first publication |  |   |
| Cenelec            | EN 45502-2-3:2010<br>Active implantable medical devices — Part 2-3: Particular requirements for cochlear and auditory brainstem implant systems (*)  | This is the first publication |  |   |
| Cenelec            | EN 60601-1:1990<br>Medical electrical equipment — Part 1: General requirements for safety<br>IEC 60601-1:1988  | 23.8.1996                     |  |   |
|                    | EN 60601-1:1990/A1:1993<br>IEC 60601-1:1988/A1:1991  | 23.8.1996                     | Note 3   | The date of this publication  |
|                    | EN 60601-1:1990/A2:1995<br>IEC 60601-1:1988/A2:1995 (*)  | 23.8.1996                     | Note 3   | The date of this publication  |
|                    | EN 60601-1:1990/AC:1994  | This is the first publication |  |   |
| Cenelec            | EN 60601-1:2006<br>Medical electrical equipment — Part 1: General requirements for basic safety and essential performance<br>IEC 60601-1:2005 (*)  | 27.11.2008                    | EN 60601-1:1990 and its amendments<br>Note 2.1 | 1.6.2012  |
|                    | EN 60601-1:2006/AC:2010  | This is the first publication |  |   |

| ESO <sup>(1)</sup> | Reference and title of the harmonised standard (and reference document)  | First publication OJ          | Reference of superseded standard | Date of cessation of presumption of conformity of superseded standard<br>Note 1 |
|--------------------|--|-------------------------------|----------------------------------|---|
| Cenelec            | EN 60601-1-6:2010<br>Medical electrical equipment — Part 1-6: General requirements for basic safety and essential performance — Collateral standard: Usability<br>IEC 60601-1-6:2010 (*) | This is the first publication |                                  |   |
| Cenelec            | EN 62304:2006<br>Medical device software — Software life-cycle processes<br>IEC 62304:2006 (*)   | 27.11.2008                    |                                  |   |
|                    | EN 62304:2006/AC:2008  | This is the first publication |                                  |   |

(<sup>1</sup>) ESO: European Standards Organisation:

— CEN: Avenue Marnix 17, 1000 Bruxelles/Brussel, BELGIQUE/BELGIË, Tel. +32 25500811; Fax +32 25500819 (<http://www.cen.eu>),  
 — Cenelec: Avenue Marnix 17, 1000 Bruxelles/Brussel, BELGIQUE/BELGIË, Tel. +32 25196871; Fax +32 25196919 (<http://www.cenelec.eu>),  
 — ETSI: 650 route des Lucioles, 06921 Sophia Antipolis, FRANCE, Tel. +33 492944200; Fax +33 493654716 (<http://www.etsi.eu>).

(\*) This European Standard does not necessarily cover the requirements introduced by Directive 2007/47/EC.

Note 1: Generally the date of cessation of presumption of conformity will be the date of withdrawal (dow), set by the European Standardisation Organisation, but attention of users of these standards is drawn to the fact that in certain exceptional cases this can be otherwise.

Note 2.1: The new (or amended) standard has the same scope as the superseded standard. On the date stated, the superseded standard ceases to give presumption of conformity with the essential requirements of the directive.

Note 2.2: The new standard has a broader scope than the superseded standard. On the date stated the superseded standard ceases to give presumption of conformity with the essential requirements of the directive.

Note 2.3: The new standard has a narrower scope than the superseded standard. On the date stated the (partially) superseded standard ceases to give presumption of conformity with the essential requirements of the directive for those products that fall within the scope of the new standard. Presumption of conformity with the essential requirements of the directive for products that still fall within the scope of the (partially) superseded standard, but that do not fall within the scope of the new standard, is unaffected.

Note 3: In case of amendments, the referenced standard is EN CCCCC:YYYY, its previous amendments, if any, and the new, quoted amendment. The superseded standard (column 3) therefore consists of EN CCCCC:YYYY and its previous amendments, if any, but without the new quoted amendment. On the date stated, the superseded standard ceases to give presumption of conformity with the essential requirements of the directive.

#### NOTE:

- Any information concerning the availability of the standards can be obtained either from the European Standardisation Organisations or from the national standardisation bodies of which the list is annexed to the Directive 98/34/EC of the European Parliament and of the Council <sup>(1)</sup> amended by the Directive 98/48/EC <sup>(2)</sup>.

<sup>(1)</sup> OJ L 204, 21.7.1998, p. 37.

<sup>(2)</sup> OJ L 217, 5.8.1998, p. 18.

- Harmonised standards are adopted by the European Standardisation Organisations in English (CEN and Cenelec also publish in French and German). Subsequently, the titles of the harmonised standards are translated into all other required official languages of the European Union by the National Standards Bodies. The European Commission is not responsible for the correctness of the titles which have been presented for publication in the Official Journal.
  - Publication of the references in the *Official Journal of the European Union* does not imply that the standards are available in all the Community languages.
  - This list replaces all the previous lists published in the *Official Journal of the European Union*. The Commission ensures the updating of this list.
  - More information about harmonised standards is available on the Internet at:  
[http://ec.europa.eu/enterprise/policies/european-standards/harmonised-standards/index\\_en.htm](http://ec.europa.eu/enterprise/policies/european-standards/harmonised-standards/index_en.htm)
-